

## CERTIFICATE OF ANALYSIS

### ANKOM Fiber Check Sample

**Lot#: ALF-2024-001**

**Alfalfa Check Sample (50g)**

Expiration Date: December 31, 2027

**Crude Fiber: 29.5%, SD: 1.55**

**ADF: 35.13%, SD: 0.69**

**NDF: 40.65%, SD 1.00**

The above sample was tested “**as received**” for the above analyses in the laboratory at ANKOM Technology using ANKOM Delta.

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.