



CERTIFICATE OF ANALYSIS

ANKOM Fat Check Sample

Lot#: OM-2024-001

Check Sample (50g)

Expiration Date: December 31, 2026

Crude Fat: 3.14%, SD: 0.16

Moisture: 6.20%, SD: 0.09

Total Fat: 5.19%, SD: 0.30

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.