



Ankom Technology
2052 O'Neil Road
Macedon, NY 14502
(315) 986-8090

CERTIFICATE OF ANALYSIS

Chemical Name: Ankom Cruder Fiber Acid, 0.255N


Ankom Part #: FCFA20CB

CAS No.: Mixture

LOT AD-24264-5

Expiration Date: September 2027

	<u>SPECIFICATION</u>	<u>RESULT</u>
Appearance	Clear to slightly cloudy liquid	conforms
Titration using 0.1M NaOH	0.244 – 0.264 N	0.252 N

Approved By: 
Samantha Jackson

Date: 10/2/2024

